

NOTES :

GENERAL

OUTPUT TO THE INSTRUMENT FROM THE PREAMP IS A GAIN OF 1.


WHEN A COMBINATION ELECTRODE IS USED, REFERENCE CONNECTION IS MADE THROUGH THE BNC CONNECTOR. USE OF A SEPARATE REFERENCE REQUIRES THAT ITS LEAD BE CONNECTED DIRECTLY TO THE Iso Grd TERMINAL IN THE INSTRUMENT.

TEMPERATURE COMPENSATION


STANDARD 8710 TC IS ACCOMPLISHED WITH A 3K Ω BALCO ELEMENT ACROSS T+ AND T-.

SHIELDING

TAKE CARE NOT TO ALLOW THE BNC SHIELD (REFERENCE LEAD) TO CONTACT METAL PIPING OR J-BOXES AT ANY POINT, AS THIS WILL SHORT THE REFERENCE CELL TO EARTH GROUND, AND CAUSE UNSTABLE READINGS. SHIELDING OF THE CABLE FROM AMPLIFIER TO INSTRUMENT IS USUALLY NOT REQUIRED. IF A SHIELD IS USED, CONNECT IT TO THE SHLD TERMINAL, AND MAKE SURE THAT IT DOES NOT TOUCH ANY OF THE OTHER LEADS, OR EARTH GROUND.



NOTES UNLESS OTHERWISE SPECIFIED:

 11751 Markon Drive Garden Grove, CA. 92841 USA phone: 714-895-4344 fax: 714-894-4839					
				DRAWN 2/21/95 rh TITLE SSA2 WITH SIGNET 8710	
CHECKED		SCALE NONE		DRAWING NO.	REV.
APPROVED		SHEET 1 OF 1	DRAWING NO. SSA2-8710		/